

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
<small>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.</small>				
1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE Dec 95		3. REPORT TYPE AND DATES COVERED Final 1 Jan 91 - 31 Aug 95
4. TITLE AND SUBTITLE Near Field Optical Microscopy Using a Vibrating Knife Edge or Stylus			5. FUNDING NUMBERS DAAL03-91-G-0014	
6. AUTHOR(S) Adrian Korpel				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) University of Iowa Iowa City, IA 52240			8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Research Office P.O. Box 12211 Research Triangle Park, NC 27709-2211			10. SPONSORING/MONITORING AGENCY REPORT NUMBER ARO 27591.2-PH	
11. SUPPLEMENTARY NOTES The view, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy, or decision, unless so designated by other documentation.				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited			12b. DISTRIBUTION CODE	
13. ABSTRACT (Maximum 200 words) The problem studied was the detection of sub-wavelength detail by scanning near-field microscopy, using a vibrating knife-edge/stylus rather than an aperture probe. A related problem was phase detection of an optical field by such scanning, not necessarily in the sub-wavelength region. (continued on reverse side)				
14. SUBJECT TERMS			15. NUMBER OF PAGES	
16. PRICE CODE				
17. SECURITY CLASSIFICATION OF REPORT UNCLASSIFIED		18. SECURITY CLASSIFICATION OF THIS PAGE UNCLASSIFIED		19. SECURITY CLASSIFICATION OF ABSTRACT UNCLASSIFIED
				20. LIMITATION OF ABSTRACT UL

MOST IMPORTANT RESULTS

1. Demonstration of one-dimensional principle of operation.
2. Demonstration of non sub-wave length two-dimensional microscope with vibrating knife edge corner.
3. Demonstration of non sub-wave length two-dimensional tomographic microscope with vibrating knife edge.
4. Analysis and simulation of evanescent wave to plane wave conversion.
5. Analysis and demonstration of phase measuring properties of one-dimensional vibrating knife edge microscope.

FINAL TECHNICAL REPORT

Grant DAAL 03-91- G-0014
period 1/01/91-8/31/95

Near Field Optical Microscopy Using a Vibrating Knife Edge or Stylus

Principal Investigator: Adrian Korpel, University of Iowa

PROBLEM STUDIED

The problem studied was the detection of sub-wavelength detail by scanning near-field microscopy, using a vibrating knife-edge/stylus rather than an aperture probe. A related problem was phase detection of an optical field by such scanning, not necessarily in the sub-wavelength region.

MOST IMPORTANT RESULTS

1. Demonstration of one-dimensional principle of operation.
2. Demonstration of non sub-wave length two-dimensional microscope with vibrating knife edge corner.
3. Demonstration of non sub-wave length two-dimensional tomographic microscope with vibrating knife edge.
4. Analysis and simulation of evanescent wave to plane wave conversion.
5. Analysis and demonstration of phase measuring properties of one-dimensional vibrating knife edge microscope.

PUBLICATIONS

1. Korpel, A., Samson, C. and Feldbush, K., " Two-dimensional operation of a scanning optical microscope using a knife edge corner.", Int. Jnl. of Im. Syst. and Technol., 4, 207-213, 1992.
2. Korpel, A., Samson, C. and Feldbush, K., "Progress in vibrating stylus near field optical microscopy. ", Proc. Workshop on Near Field Optics, Arc et Senans, France, October 1992, p. 399-406

3. Samson.A and Korpel. A, "Two- dimensional operation of a scanning microscope by vibrating knife edge tomography", App. Opt. 95, 285-289, 1995
4. S.Samson, A.Korpel and H.S.Snyder, "Conversion of propagating waves by vibrating knife edge,' Int. Jnl. Im. Syst. and Technology , in press.
5. Adrian Korpel and Holly Snyder "Coherence properties of vibrating knife edge field sampling." In preparation.

Personnel supported:

Adrian Korpel (PI)

Scott Samson (graduate student) PhDEE

Kurt Feldbush (under an AASERT award) MSEE

Holly Snyder (under an AASERT award) MSEE